



THE UNIVERSITY *of* EDINBURGH

Edinburgh Research Explorer

Micro-CT based finite element models of cancellous bone predict accurately displacement once the boundary condition is well replicated: A validation study

Citation for published version:

Chen, Y, Dall'Ara, E, Sales, E, Manda, K, Wallace, R, Pankaj, P & Viceconti, M 2017, 'Micro-CT based finite element models of cancellous bone predict accurately displacement once the boundary condition is well replicated: A validation study', *Journal of the mechanical behavior of biomedical materials*, vol. 65, pp. 644–651. <https://doi.org/10.1016/j.jmbbm.2016.09.014>

Digital Object Identifier (DOI):

[10.1016/j.jmbbm.2016.09.014](https://doi.org/10.1016/j.jmbbm.2016.09.014)

Link:

[Link to publication record in Edinburgh Research Explorer](#)

Document Version:

Publisher's PDF, also known as Version of record

Published In:

Journal of the mechanical behavior of biomedical materials

Publisher Rights Statement:

Under a Creative Commons license


General rights

Copyright for the publications made accessible via the Edinburgh Research Explorer is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The University of Edinburgh has made every reasonable effort to ensure that Edinburgh Research Explorer content complies with UK legislation. If you believe that the public display of this file breaches copyright please contact openaccess@ed.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.





Short communication

Micro-CT based finite element models of cancellous bone predict accurately displacement once the boundary condition is well replicated: A validation study



Yuan Chen^a, Enrico Dall'Ara^b, Erika Sales^c, Krishnagoud Manda^c, Robert Wallace^d,
Pankaj Pankaj^c, Marco Viceconti^{a,*}

^a Department of Mechanical Engineering and INSIGNEO Institute for in silico Medicine, University of Sheffield, UK

^b Department of Human Metabolism and INSIGNEO Institute for in silico Medicine, University of Sheffield, UK

^c Institute for Bioengineering, School of Engineering, The University of Edinburgh, UK

^d Department of Orthopaedics, The University of Edinburgh, UK

ARTICLE INFO

Keywords:

MicroCT
microFE
Cancellous bone
DVC
Validation

ABSTRACT

Non-destructive 3D micro-computed tomography (microCT) based finite element (microFE) models are used to estimate bone mechanical properties at tissue level. However, their validation remains challenging. Recent improvements in the quantification of displacements in bone tissue biopsies subjected to staged compression, using refined Digital Volume Correlation (DVC) techniques, now provide a full field displacement information accurate enough to be used for microFE validation. In this study, three specimens (two humans and one bovine) were tested with two different experimental set-ups, and the resulting data processed with the same DVC algorithm. The resulting displacement vector field was compared to that predicted by microFE models solved with three different boundary conditions (BC): nominal force resultant, nominal displacement resultant, distributed displacement. The first two conditions were obtained directly from the measurements provided by the experimental jigs, whereas in the third case the displacement field measured by the DVC in the top and bottom layer of the specimen was applied. Results show excellent relationship between the numerical predictions (x) and the experiments (y) when using BC derived from the DVC measurements (U_x : $y=1.07x-0.002$, RMSE: 0.001 mm; U_y : $y=1.03x-0.001$, RMSE: 0.001 mm; U_z : $y=x+0.0002$, RMSE: 0.001 mm for bovine specimen), whereas only poor correlation was found using BCs according to experiment set-ups. In conclusion, microFE models were found to predict accurately the vectorial displacement field using interpolated displacement boundary condition from DVC measurement.

1. Introduction

Bone tissue is a complex hierarchical material (Cowin, 2001). In order to address clinical and preclinical problems, it is important to study bone at the spatial scale that allows the most appropriate characterization of its mechanical behaviour. At the tissue scale, the interaction between bone mechanical stimuli and the biological function driven by the cell activity becomes more evident (Viceconti, 2012). The microCT based finite element (microFE) method has become a popular tool for non-destructive structural analysis of cancellous bone tissue (Hollister et al., 1994; van Rietbergen et al., 1995; Verhulst et al., 2008). The method involves the direct conversion of the 3D voxels of micro-computed tomography (microCT) images of the bone tissue (Feldkamp et al., 1989) into equally shaped and sized hexahedral

elements. As microCT imaging has the ability to accurately resolve bone morphology in great detail (Bouxsein et al., 2010), specimen-specific microFE models that represent the structure of the specimen can be generated (Ulrich et al., 1998).

Every modelling method requires, before it can be considered reliable, a complete verification, validation, and uncertainty quantification assessment (Anderson et al., 2007). A systematic verification analysis of microFE models of bone tissue was recently published (Chen et al., 2014). However, for validation, the number of published reports is limited. While the predicted apparent properties (e.g. stiffness, strength) of each specimen can be compared with accurate experimental measurements (Christen et al., 2013; Pistoia et al., 2002; Wolfram et al., 2010; Yeni and Fyhrie, 2001), the validation of such models for local predictions is not trivial.

* Correspondence to : INSIGNEO Institute for in silico Medicine Engineering Graduate School The Pam Liversidge Building Mappin Street Sheffield S1 3JD.
E-mail address: m.viceconti@shef.ac.uk (M. Viceconti).

<http://dx.doi.org/10.1016/j.jmbbm.2016.09.014>

Received 19 September 2015; Received in revised form 13 July 2016; Accepted 9 September 2016

Available online 17 September 2016

1751-6161/© 2016 The Authors. Published by Elsevier Ltd. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

One possible approach is to use a full-field method, such as Digital Volume Correlation (DVC) techniques, to extract displacement or strain fields from repeated microCT scans performed during stepwise-compression experiments. In every validation study the boundary conditions (BCs) imposed in the model should be the same as in the experiments. However, even if we can measure accurately the resultant force applied, or the total displacement imposed during the stepwise compression test, the aspect ratio of the specimens typically used in these tests might be too small to assume valid Saint-Venant's Principle. If this is the case, it is not enough to reproduce in the model the loading resultant, but we need to consider also how such forces are locally distributed. To the authors' knowledge there are two studies in the literature that used DVC measurements to validate microFE models displacement predictions on cancellous bone specimens, and their results are somehow inconclusive. Recently, [Zhu et al. \(2015\)](#) compared the predictions of microCT based tetrahedral homogeneous models to DVC measurements for bovine bone interdigitated with acrylic cement and a cellular foam samples. Only qualitative comparison between models and DVC displacement only along the loading direction was reported. [Zaue et al. \(2006\)](#) was the first to use a DVC approach based on the one reported in [Bay et al. \(1999\)](#) to quantitatively validate a linear elastic microFE model of cancellous bone. They found very good correlation in displacement measured along the major loading direction (R^2 from 0.91 to 0.97, slopes between 0.93 and 0.98), but only poor correlation for transverse displacements (R^2 from 0.29 to 0.60, slopes between 0.33 and 0.88). This result is surprising as the precision error of their DVC method is isotropic ([Dall'Ara et al., 2014; Zaue et al., 2006](#)) and the predictions of the microFE models should not be affected by the loading direction unless strong local anisotropy needs to be included in the models. The findings reported by [Zaue et al. \(2006\)](#) suggest that homogeneous isotropic microFE models is not reliable in predicting transverse displacement. Therefore, we need to further explore the ability of predicting local displacement from microFE models, widely used in the research community to estimate bone properties at the tissue levels.

The aim of this study was to determine the accuracy of microFE models, generated from microCT data, in predicting the displacements of cancellous bone specimens subjected to compression, when sufficiently accurate full-field displacement measurements are used for the validation, and appropriate boundary conditions are simulated.

2. Materials and methods

Two independent testing procedures were used in order to assess the sensitivity of the validation approach for two different experimental protocols and input images and to extend the validity of the results. In both cases similar workflows ([Fig. 1](#)) were used.

2.1. Specimen preparation and scanning

All procedures on human tissue were performed with approval from the Research Ethics Committee for use of discarded bone material (LREC 2002/1/22).

Two cylindrical specimens (Specimen1: height equal to 13.2 mm, diameter equal to 10.6 mm; Specimen2: height equal to 11.5 mm, diameter equal to 10.6 mm) were extracted from the central part of two human femoral heads from patients who underwent total hip replacement using diamond-tipped cores (Starlite Industries, Rosemount PA, USA). Specimen 1 was from an osteoarthritic male aged 68 and Specimen 2 from a 94 years old male. The ends of the core samples were cut parallel using a Buehler Isomet low speed saw (Buehler, Illinois, USA).

The third cylindrical specimen (Specimen3: height equal to 11.88 mm, diameter equal to 7.89 mm) was drilled (diamond core drill with nominal internal diameter equal to 8 mm mounted on a pillar drilling machine, GDM50B, Sealey, UK) from a bone slice cut (0.2 mm




Fig. 1. Workflow of the study.

diamond bandsaw mounted on a 300 CP, Exakt GmbH, Germany) from a bovine femoral greater trochanter (female, 18 months old). All operations were performed under constant water irrigation in order to reduce potential damage to the bone specimen. Animal tissue was extracted from a bovine femur, collected from an animal culled for alimentary purposes.

The specimens were then microCT scanned (Skyscan 1172; Specimen1 and Specimen2: voxel size 17.22 μm , 54 kV, 185 μA , 0.5 mm aluminium filter, exposure time 885 ms, no averaging; Specimen3: voxel size 9.92 μm , 59 kV, 169 μA , 1 mm aluminium filter, exposure time 1180 ms, averaged by two frames). Each image was cropped in order to include only the bone specimens and datasets were subsampled by a factor of two (ImageJ, V1.50a), resulting in a new voxel size equal to 34.44 μm and 19.84 μm for human and bovine specimens, respectively. For the Specimen1 and Specimen2 top and bottom slices with partial bone and air were removed, while for Specimen3 slices in the embedding material were removed. Bone volume fraction (BV/TV), trabecular thickness (Tb.Th), trabecular spacing (Tb.Sp), degree of anisotropy (DA) and angle between the main trabecular direction and the loading axis (α_Z) were computed with the ImageJ plugin BoneJ ([Doubé et al., 2010](#)). Information for all specimens is summarized in [Table 1](#).

2.2. In situ mechanical testing

All three specimens were mechanically tested within the microCT system.

Specimen1 and Specimen2 were fully hydrated, and tested with the compressive device provided by the manufacturer of the microCT with a 440 N load cell. The specimens were positioned in between two parallel loading plates, in the middle of the device. A first scan (undeformed) was performed with the specimens under a small preload of 7.22 N in order to avoid motion artefacts. Afterwards, a compressive step up to 1% apparent strain was applied without repositioning and the specimen was scanned in its deformed configuration ([Fig. 2, left](#)).

Specimen3 was tested in a custom-made compressive device to be positioned within the same microCT model. The load was applied by a manual screw-ball joint mechanism and was measured with a 2 kN load cell (LPM530, Cooper Instruments & Systems, Warrenton, USA). The 1.5 mm external portions of the specimen were embedded in PMMA (Technovit 4071, Heraeus Kulzer GmbH, Wehrheim, Germany)

Table 1
Specimens' information.

Specimen	Species	Location	Height (mm)	Diameter (mm)	BV/TV (%)	Tb.Th (μm)	Tb.Sp (μm)	DA (dimensionless)	α.Z (deg)	Voxel size (μm)	Force (N)	Disp (mm)
Specimen1	Human	Femoral head	13.2	10.6	30.29	192 ± 69	427 ± 193	0.571	27	17.22	42	0.13
Specimen2	Human	Femoral head	11.5	10.6	29.64	188 ± 67	376 ± 160	0.594	9	17.22	162	0.12
Specimen3	Bovine	Greater trochanter	11.88	7.89	22.82	171 ± 51	550 ± 152	0.539	60	9.92	120	Na

Na: no displacement was measured and this type of BCs was not modelled.




Fig. 2. Scheme of the loading jigs used for testing Specimen1 and Specimen2 (left) and Specimen3 (right).




Fig. 3. 3D representation of microCT (left) of Specimen3 and the corresponding microFE model (right).

after proper alignment with the loading axis of the jig. A first scan (undeformed) was performed with the specimen under a preload of 2 N in order to avoid motion artefacts. Afterwards, a compressive step to 120 N was applied without repositioning and the specimen was scanned in its deformed configuration. A liquid-filled chamber was used in order to keep the specimen submerged in a 0.9% NaCl solution during the test (Fig. 2, right).

2.3. DVC measurement of displacement

The DVC method computes the field of displacements by registering elastically the couple of undeformed and deformed images for each specimen (34.44 μm and 19.84 μm). In the present study, we used an in-house elastic registration library (Sheffield Image Registration Toolkit, ShIRT) (Barber and Hose, 2005; Barber et al., 2007; Khodabakhshi et al., 2013). The registration equations are solved in the nodes of a grid superimposed to both images to be registered and with certain nodal spacing (NS), assuming a linear behaviour in

displacement in between the nodes. In the current study, we used NS equal to 12 voxels (~413 μm) for human cancellous bone (Specimen1 and Specimen2) and NS 25 voxels (~496 μm) for bovine cancellous bone (Specimen3). With this NS the accuracy and precision in displacement is approximately $0.00016 \pm 0.00034 \mu\text{m}$ (~400 μm) for Specimen1 and Specimen2 and $0.0000098 \pm 0.00014 \mu\text{m}$ (~500 μm) for Specimen3 (Palanca et al., 2015). As for all other methods described in the literature, also the accuracy of our method was calculated by using virtually moved images; it should be noted that this test provides an estimate of the lower boundary of the error. Still, the method used here is significantly more accurate than any other described previously.

2.4. microFE models

To reduce the computational cost, the original image datasets were first subsampled by a factor of two as for the DVC measurements (34.44 μm for Specimen1 and Specimen2; 19.84 μm for Specimen3). All image datasets were filtered by a Gaussian filter with $\sigma=1.2$ and a support of 3 voxels to reduce the high frequency noise (Christen et al., 2014). Then the images were binarized using a single-level threshold by finding the mean value between two peaks (one representing the bone tissue, one representing the background) in the grayscale histograms. Voxels below the threshold value were deleted and for those above the threshold value, a connectivity filter (Matlab, R2014b, Mathworks, Inc.) was applied to remove the isolated voxels. In particular, only elements with surface (four nodes) connectivity were kept in the model. Finally, each remaining voxel in the image datasets was converted directly into equally sized 8-node hexahedral elements (Fig. 3). The material properties for all microFE models were assumed to be linear and isotropic, with a uniform Young's modulus of 17 GPa (Bayraktar et al., 2004; Morgan et al., 2003) and a Poisson's ratio of 0.3 (Pistoia et al., 2002). Following the procedure by Chen et al. (2014), a convergence study was performed on models with three BCs (mentioned in Section 2.5). All models (element size equal to 34.44 μm for Specimen1 and Specimen2 and 19.84 μm for Specimen3) reached a local convergence in displacement, with a difference of less than 1% compared to models of the most mesh refined microFEs with element size equal to the original microCT voxel size (17.22 μm for Specimen1 and Specimen2 and to 9.92 μm for Specimen3).

2.5. Boundary conditions

Three different BCs for each microFE simulation were used. The first set of BCs (hereinafter referred to as “force BCs” was defined based on the experimentally measured force. The layer of nodes of the microFE belonging to the fixed surface of the bone were fixed in the loading direction and a vertical force was distributed equally on each node of the loaded surface so that their resultant was equal to the measured axial force (42N for Specimen1, 162N for Specimen2 and 120N for Specimen3). The nodes of the both loaded and fixed surface were free to move in transverse direction for the Specimen1 and

Table 2

Correlation between microFE prediction and DVC measurement.

Specimen	BCs	Direction	Number of points (Remaining percentage)	R ²	Slope	Intercept (mm)	ICC	RMSE (mm)	RMSE%	Max.err (mm)	Max.err%
Specimen1	Force	U _x	3920(97%)	0.77	33.81	−0.040	0.02	0.036	54.1%	0.066	99.3%
		U _y	3923(97%)	0.55	15.48	0.050	0.02	0.035	61.3%	0.057	100.1%
		U _z	3930(97%)	0.47	11.41	0.100	Na	0.082	80.3%	0.103	101.4%
	Disp	U _x	3951(98%)	0.06	0.30	−0.030	0.18	0.027	38.7%	0.070	92.6%
		U _y	3915(97%)	0.41	0.43	0.050	0.08	0.078	136.6%	0.099	173.8%
		U _z	3947(98%)	0.02	0.05	0.080	0.01	0.154	152.2%	0.207	203.8%
	Intep	U _x	3890(96%)	0.99	1.00	< 0.001	0.99	0.001	1.7%	0.005	7.4%
		U _y	3888(96%)	0.99	1.01	< 0.001	0.99	0.001	1.7%	0.004	7.1%
		U _z	3916(97%)	0.99	0.99	< 0.001	0.99	0.001	1.1%	0.004	3.5%
Specimen2	Force	U _x	3612(98%)	0.28	−29.18	0.110	Na	0.103	61.1%	0.169	100.0%
		U _y	3603(98%)	0.27	−52.88	−0.210	Na	0.167	58.7%	0.284	99.9%
		U _z	3619(99%)	< 0.01	−3.04	0.280	Na	0.308	70.1%	0.440	101.2%
	Disp	U _x	3633(99%)	0.05	0.93	0.100	0.02	0.104	61.7%	0.171	101.2%
		U _y	3638(99%)	0.52	−3.61	−0.240	Na	0.149	52.6%	0.291	102.2%
		U _z	3619(99%)	< 0.01	−0.21	0.280	Na	0.363	82.3%	0.537	122.0%
	Intep	U _x	3511(96%)	0.99	1.03	−0.006	0.99	0.003	2.0%	0.018	10.7%
		U _y	3516(96%)	0.99	1.02	0.006	0.99	0.005	1.6%	0.015	5.2%
		U _z	3531(96%)	0.99	0.98	−0.005	0.99	0.011	2.4%	0.025	5.6%
Specimen3	Force	U _x	586(99%)	0.07	1.48	0.030	0.02	0.035	58.0%	0.060	98.3%
		U _y	574(97%)	0.03	1.13	0.030	0.01	0.028	58.2%	0.049	100.5%
		U _z	577(98%)	0.32	1.55	−0.003	0.42	0.016	28.7%	0.039	72.1%
	Intep	U _x	570(97%)	0.99	1.07	−0.002	0.99	0.001	2.1%	0.003	5.6%
		U _y	573(97%)	0.97	1.03	−0.001	0.99	0.001	2.4%	0.003	7.1%
		U _z	568(96%)	0.99	1.00	< 0.001	0.99	0.001	1.5%	0.002	4.5%

*Correlations are all significant at $p < 0.01$; Na: ICC not reliable.

Specimen2 (free boundary conditions) and fixed in the transverse directions for Specimen3 (simulation of embedding).

The second set of BCs (later referred to as “displacement BCs”) was defined based on the experimentally measured displacements. The layer of nodes of the microFE belonging to the fixed surface of the bone were fixed in only the loading direction and the experimentally measured vertical displacement was imposed to each node of the loaded surface (130 μm for Specimen1, 115 μm for Specimen2; for Specimen3 no displacement was measured and this type of BCs was not modelled), leaving free to move in the transverse directions.

The third set of BCs (later referred to as “interpolated BCs”) was imposed from the DVC measurements. The top and bottom layers of nodes from the microFE were superimposed into the displacement field measured from the DVC approach. The displacements of the boundary nodes were assigned by linear interpolation using element shape functions.

The resulting models were composed of approximately 8 million elements for Specimen1 and Specimen2 and 11.5 million for Specimen3. All simulations were performed using ANSYS (Release 15.0, ANSYS, Inc.) on a high performance parallel computing cluster (Iceberg, <https://www.shef.ac.uk/wrgrid/iceberg>) with 32 cores (CPU: Intel (R) Xeon E-5-2670. 2.60 GHz, RAM: 32 GB/core). The computational time for linear elastic simulation ranged from 8–10 min (for Specimen1 and Specimen2) to 13–15 minutes (Specimen3) using ANSYS preconditioned conjugate gradient (PCG) solver (tolerance 1×10^{-08}).

2.6. Comparison between experimental and computational results

The comparison was limited to the middle 80% of the specimen to avoid boundary effect. The DVC procedure provides the displacements at the nodes of its grid (with NS 12 or 25 voxels according to the specimen) over the whole image. Due to way the DVC grid and of the microFE models are generated, each node of the DVC grid lays either in the centroid of a microFE element or in the marrow. The comparison of the results was performed only in the locations of the DVC nodes

corresponding to microFE elements (i.e. in the bone). For those locations, the displacements measured with the DVC were compared to the interpolated value in the centroid of the element of the microFE. Following this procedure, we obtained the following number of comparison pairs: 4041 for Specimen1, 3671 for Specimen2 and 589 for Specimen3. In each specimen, we compared the displacements predicted by microFE under the different BCs to the DVC results.

2.7. Statistics

Any observation with Cook's distance (Fox and Long, 1990) larger than five times the mean Cook's distance was considered as outliers and removed from the analysis. This approach removed 1% to 4% points for each analysis. The comparison of displacement for microFE models and experiments was performed using linear regression, where the slope and intercept of the equation as well as coefficient of determination (R^2) were reported. For each comparison the Root mean square error (RMSE), the RMSE divided by the maximum experiment value (RMSE%), the largest difference between microFE prediction and DVC measurements (Max.error), the Max.error divided by the maximum experimental value (Max.error%), and the intra-class correlation (ICC) were computed. The number of comparison pairs remained in each analysis was reported in Table 2.

3. Results

All coefficients calculated from the correlations between predicted and measured displacements are reported in Table 2.

The displacement predicted by microFE models with three different BCs and DVC measurements were all significantly linearly correlated ($p < 0.01$). However, microFE models with “force BCs” and “displacement BCs” were far from the 1:1 relationship, underlined by the low ICCs (from 0.02 to 0.42). Conversely, microFE models with “interpolated BCs” lead to excellent correlations, with slope close to one (range: 0.98 to 1.07), intercept close to zero (range: −0.006 to 0.006 mm), high R^2 (range: 0.97 to 0.99) and high ICC (0.99). In that




Fig. 4. Displacement predicted by microFE plotted against DVC measurement using “interpolated BC” for Specimen1 (top), Specimen2 (middle) and Specimen3 (bottom).

case similar results were found for the three specimens (Fig. 4), with RMSE% lower than 2.5% (with maximum equal to 2.4% for predictions of U_z for Specimen2 and of U_y for Specimen3) and Max.err% lower than 11% (with maximum equal to 10.7% for predictions of U_x for Specimen2). These models overall predicted better U_z (displacements along the major compression direction) with RMSE% from 1.1% to 2.4% and Max.err% from 3.5% to 5.6% compared to the displacements along the transverse directions (RMSE% from 1.7% to 2.1% and Max.err% from 5.6% to 10.7% for U_x and RMSE% from 1.7% to 2.4% and Max.err% from 5.2% to 7.1% for U_y). The best correlation was found for predictions of U_z for Specimen3, with slope equal to 1 and intercept equal to 0.0002 mm.

In Fig. 5 comparisons between the predicted and measured vertical displacement for the “interpolated BCs” are reported. From those graph it can be noted that all specimens have rotated to some extent during the experiments, especially Specimen1 and Specimen2.

The scatter plots for all three specimens using “force BCs” and “displacement BCs” can be found in the Supplementary materials (available through the FigShare on-line <https://figshare.com/s/0fa91152b09dfe5c5142>).

4. Discussion

The aim of the present study was to compare DVC measurements of displacement in cancellous bone samples with the predictions of microFE simulated with different BCs.

All microFE models using “interpolated BCs” accurately predicted the displacement field measured experimentally, as induced by loads small enough to assume linear elasticity, with similar accuracy in all spatial directions. It should be noted that similar results were found over experiments generated by two separate research groups, using different experimental protocols, and imaged at different resolutions, underlying the robustness of the method.

On the other hand, microFE models using “force BCs” and “displacement BCs” in general predicted displacement poorly, when compared to DVC measurements. These poor predictions could be due to the fact that the friction between loading plates and surface of the sample (which in some studies is artificially reduced, but that can never be nullified) was not modelled (for “force BCs”), or by the potential parallelism error between the two flat surfaces of the specimens. (“displacement BCs”, for Specimen1 and Specimen2 which were not embedded) and/or by the potential micro-movements of the loading




Fig. 5. Contour plot of the displacement along the loading direction Z viewed in the coronal plane for Specimen1 (top), Specimen2 (middle) and Specimen3 (bottom). Predictions from the microFEs with “interpolated BCs” (left) and DVC measurements (right) are reported.

plate (“displacement BCs”, all three specimens). This effect does not seem to be related to the different inclination between the main trabecular direction of each sample and the loading direction (from

9° to 60°), at least for relatively dense samples as those considered in this study ($23\% < BV/TV < 30\%$).

The excellent correlations found for displacements along the

loading direction (R^2 equal to 0.99, slope from 0.98 to 1.00, intercept from -5 to -0.2 μm , ICC equal to 0.99) is comparable to what was reported by Zauel et al. (2006) on a similar study performed on two human cancellous specimens (R^2 from 0.91 to 0.97; slope from 0.93 to 0.98; intercept from 79 to 145 μm). However, in that case much worse predictions were found for the transverse directions (R^2 from 0.29 to 0.60; slope from 0.33 to 0.88; intercept from -954 to 40 μm) while in the present study they were excellent, even if with slightly larger scatter (R^2 from 0.97 to 0.99; slope from 1.00 to 1.07; intercept from -6 to 6 μm). The small differences between the predictions of the displacements along the different directions found in our study underline that the assumption of isotropic material property for every element in case of microFE models is well posed. The improvement in the predictive accuracy of the displacements along the transverse directions in this study compared to the previous report (Zauel et al., 2006) may be due to the improved accuracy of the DVC method used in this work (Dall'Ara et al., 2014; Palanca et al., 2015). It remains to be investigated how sensitive the models are for strain predictions.

The main limitation of this study is that it validates microFE models only with respect to displacements, whereas in many studies these models are primarily used to predict strains. However, the accuracy and precision of the current experimental methods for strain measurement in each element (as small as 10–20 μm) when modelling whole bone biopsy (10–20 mm large) are too low (Grassi and Isaksson, 2015; Palanca et al., 2015). All available DVC methods can provide reasonably accurate quantifications of strains only over much coarser resolutions (i.e. 40–50 voxels); while in principle we could sub-sample the microFE model at the same lower resolution, this comparison would be meaningless, as we would need to assume bone as a continuum. Such validation has been already done at the organ level by using strain-gauges (Cristofolini and Viceconti, 2000; Cristofolini et al., 1996; Taddei et al., 2010; Viceconti et al., 2001) or digital image correlation (Grassi et al., 2016). Further research is required to produce a controlled experiment where the strain field can be measured with sufficient accuracy with a special resolution of 10–20 μm . Here we investigated the simplest type of microFE models, based on Cartesian homogenous elements. It might be interesting to explore if the predictive accuracy is further improved by adding local heterogeneity (Gross et al., 2012), or local smoothing (Muller and Rueggsegger, 1995) to overcome the inevitable partial volume effect induced by homogeneous models. Moreover, if larger compressive loads would be analysed, nonlinear models (Harrison et al., 2013) would become fundamental in order to simulate the local yielding of the trabecular bone structure. Finally, more specimens should be tested in order to increase the applicability of the model, possibly with increased aspect ratio in order to reduce the effect of the loading plates on the displacement field. All these improvements and further application to different bone microstructures are currently under development.

In conclusion, microFE models with “interpolated BCs” predict local displacements in cancellous bone samples with excellent accuracy in all spatial directions. Having provided consistent results across all three specimens tested by different experimental set-ups, we can conclude microFE models are robust once the BC assigned is well matched with experiments and, therefore, can be used as a modelling framework in the future to study bone mechanics at the tissue level, at least to predict displacement. It remains to be investigated whether this hold valid for the local strain and if improvement in the models (e.g. by adding local heterogeneity, smooth surfaces, nonlinearities) would lead to more accurate predictions.

Acknowledgements

The authors would like to thank the HPC Iceberg team at University of Sheffield for providing computational resources, the contribution of Anil Kumar for his kind advice on structural analysis with ANSYS, Mr K Rotchell and Mr J Bradley for the help with manufacturing the jig

and Prof D Barber for sharing the elastic registration toolkit.

This study was partially supported by EPSRC funded project MultiSim (Grant No. EP/K03877X/1); NC3R funded project (Grant No. 113629) and the FP7 European program MAMBO (PIEF-GA-2012-327357). The study was also partially funded by the EPSRC grant EP/K036939/1. The third author also acknowledges the support received from CNPq – Brazil.

Appendix A. Supplementary material

Supplementary data associated with this article can be found in the online version at <http://dx.doi.org/10.1016/j.jmbbm.2016.09.014>.

References

- Anderson, A.E., Ellis, B.J., Weiss, J.A., 2007. Verification, validation and sensitivity studies in computational biomechanics. *Comput. Methods Biomech. Biomed. Eng.* 10, 171–184.
- Barber, D.C., Hose, D.R., 2005. Automatic segmentation of medical images using image registration: diagnostic and simulation applications. *J. Med. Eng. Technol.* 29, 53–63.
- Barber, D.C., Oubel, E., Frangi, A.F., Hose, D.R., 2007. Efficient computational fluid dynamics mesh generation by image registration. *Med. Image Anal.* 11, 648–662.
- Bay, B.K., Smith, T.S., Fyhrie, D.P., Saad, M., 1999. Digital volume correlation: Three-dimensional strain mapping using X-ray tomography. *Experimental Mechanics*, Sept 39 (3), 217–226.
- Bayraktar, H.H., Morgan, E.F., Niebur, G.L., Morris, G.E., Wong, E.K., Keaveny, T.M., 2004. Comparison of the elastic and yield properties of human femoral trabecular and cortical bone tissue. *J. Biomech.* 37, 27–35.
- Bouxsein, M.L., Boyd, S.K., Christiansen, B.A., Guldberg, R.E., Jepsen, K.J., Muller, R., 2010. Guidelines for assessment of bone microstructure in rodents using micro-computed tomography. *J. Bone Miner. Res.* 25, 1468–1486.
- Chen, Y., Pani, M., Taddei, F., Mazza, C., Li, X., Viceconti, M., 2014. Large-scale finite element analysis of human cancellous bone tissue micro computer tomography data: a convergence study. *J. Biomech. Eng.*, Oct 136 (10), 101013. <http://dx.doi.org/10.1115/1.4028106>.
- Christen, D., Melton, L.J., 3rd, Zwahlen, A., Amin, S., Khosla, S., Muller, R., 2013. Improved fracture risk assessment based on nonlinear micro-finite element simulations from HRpQCT images at the distal radius. *J. Bone Miner. Res.* 28, 2601–2608.
- Christen, D., Zwahlen, A., Muller, R., 2014. Reproducibility for linear and nonlinear micro-finite element simulations with density derived material properties of the human radius. *J. Mech. Behav. Biomed. Mater.* 29, 500–507.
- Cowin, S.C., 2001. *Bone Mechanics Handbook* 2nd ed. CRC Press, Boca Raton, FL.
- Cristofolini, L., Viceconti, M., 2000. Mechanical validation of whole bone composite tibia models. *J. Biomech.* 33, 279–288.
- Cristofolini, L., Viceconti, M., Cappello, A., Toni, A., 1996. Mechanical validation of whole bone composite femur models. *J. Biomech.* 29, 525–535.
- Dall'Ara, E., Barber, D., Viceconti, M., 2014. About the inevitable compromise between spatial resolution and accuracy of strain measurement for bone tissue: a 3D zero-strain study. *J. Biomech.* 47, 2956–2963.
- Doube, M., Klosowski, M.M., Arganda-Carreras, I., Cordelières, F.P., Dougherty, R.P., Jackson, J.S., Schmid, B., Hutchinson, J.R., Shefelbine, S.J., 2010. BoneJ: free and extensible bone image analysis in Image. *J. Bone* 47, 1076–1079.
- Feldkamp, L.A., Goldstein, S.A., Parfitt, A.M., Jesion, G., Kleerekoper, M., 1989. The direct examination of three-dimensional bone architecture in vitro by computed tomography. *J. Bone Miner. Res.* 4, 3–11.
- Fox, J., Long, J.S., 1990. *Modern Methods of Data Analysis*. Sage Publications, Newbury Park, Calif.
- Grassi, L., Isaksson, H., 2015. Extracting accurate strain measurements in bone mechanics: a critical review of current methods. *J. Mech. Behav. Biomed. Mater.* 50, 43–54.
- Grassi, L., Vaananen, S.P., Ristinmaa, M., Jurvelin, J.S., Isaksson, H., 2016. How accurately can subject-specific finite element models predict strains and strength of human femora? Investigation using full-field measurements. *J. Biomech.* 49, 802–806.
- Gross, T., Pahr, D.H., Peyrin, F., Zysset, P.K., 2012. Mineral heterogeneity has a minor influence on the apparent elastic properties of human cancellous bone: a SRmuCT-based finite element study. *Comput. Methods Biomech. Biomed. Eng.* 15, 1137–1144.
- Harrison, N.M., McDonnell, P., Mullins, L., Wilson, N., O'Mahoney, D., McHugh, P.E., 2013. Failure modelling of trabecular bone using a non-linear combined damage and fracture voxel finite element approach. *Biomech. Model. Mechanobiol.* 12, 225–241.
- Hollister, S.J., Brennan, J.M., Kikuchi, N., 1994. A homogenization sampling procedure for calculating trabecular bone effective stiffness and tissue level stress. *J. Biomech.* 27, 433–444.
- Khodabakhshi, G., Walker, D., Scutt, A., Way, L., Cowie, R.M., Hose, D.R., 2013. Measuring three-dimensional strain distribution in tendon. *J. Microsc.* 249, 195–205.
- Morgan, E.F., Bayraktar, H.H., Keaveny, T.M., 2003. Trabecular bone modulus-density relationships depend on anatomic site. *J. Biomech.* 36, 897–904.

- Muller, R., Ruegsegger, P., 1995. Three-dimensional finite element modelling of non-invasively assessed trabecular bone structures. *Med. Eng. Phys.* 17, 126–133.
- Palanca, M., Tozzi, G., Cristofolini, L., Viceconti, M., Dall'Ara, E., 2015. Three-dimensional local measurements of bone strain and displacement: comparison of three digital volume correlation approaches. *J. Biomech. Eng.* 137.
- Pistoia, W., van Rietbergen, B., Lochmuller, E.M., Lill, C.A., Eckstein, F., Ruegsegger, P., 2002. Estimation of distal radius failure load with micro-finite element analysis models based on three-dimensional peripheral quantitative computed tomography images. *Bone* 30, 842–848.
- Taddei, F., Martelli, S., Gill, H.S., Cristofolini, L., Viceconti, M., 2010. Finite element modeling of resurfacing hip prosthesis: estimation of accuracy through experimental validation. *J. Biomech. Eng.* 132, 021002.
- Ulrich, D., van Rietbergen, B., Weinans, H., Ruegsegger, P., 1998. Finite element analysis of trabecular bone structure: a comparison of image-based meshing techniques. *J. Biomech.* 31, 1187–1192.
- van Rietbergen, B., Weinans, H., Huiskes, R., Odgaard, A., 1995. A new method to determine trabecular bone elastic properties and loading using micromechanical finite-element models. *J. Biomech.* 28, 69–81.
- Verhul, E., van Rietbergen, B., Muller, R., Huiskes, R., 2008. Indirect determination of trabecular bone effective tissue failure properties using micro-finite element simulations. *J. Biomech.* 41, 1479–1485.
- Viceconti, M., 2012. *Multiscale Modeling of the Skeletal System*. Cambridge University Press, Cambridge.
- Viceconti, M., Cristofolini, L., Baleani, M., Toni, A., 2001. Pre-clinical validation of a new partially cemented femoral prosthesis by synergetic use of numerical and experimental methods. *J. Biomech.* 34, 723–731.
- Wolfram, U., Wilke, H.J., Zysset, P.K., 2010. Valid micro finite element models of vertebral trabecular bone can be obtained using tissue properties measured with nanoindentation under wet conditions. *J. Biomech.* 43, 1731–1737.
- Yeni, Y.N., Fyhrie, D.P., 2001. Finite element calculated uniaxial apparent stiffness is a consistent predictor of uniaxial apparent strength in human vertebral cancellous bone tested with different boundary conditions. *J. Biomech.* 34, 1649–1654.
- Zauel, R., Yeni, Y.N., Bay, B.K., Dong, X.N., Fyhrie, D.P., 2006. Comparison of the linear finite element prediction of deformation and strain of human cancellous bone to 3D digital volume correlation measurements. *J. Biomech. Eng.* 128, 1–6.
- Zhu, M.L., Zhang, Q.H., Lupton, C., Tong, J., 2015. Spatial resolution and measurement uncertainty of strains in bone and bone-cement interface using digital volume correlation. *J. Mech. Behav. Biomed. Mater.* 57, 269–279.